[TAB #13]

JUL 5 2002

K021007

Attachment #11

Summary of 510(k) Submission

A. INFORMATION

1. SUBMITTER'S

NAME: **TILLOTSON HEALTHCARE CORPORATION**

ADDRESS: **360 Route 101**
Bedford, NH 03110 U.S.A.

TELEPHONE NUMBER: **(603) 472-6600**

CONTACT PERSON: **F.W. Perrella**

DATE SUMMARY PREPARED: **April 2002**

2. NAME OF DEVICE

TRADE OR PROPRIETARY NAME:
Ultra Preserve With Aloe Vera (Green) Natural Rubber Latex Examination Glove Made from Allotex® an enzyme treated natural rubber latex with a Protein Content Claim of 200 micrograms or less with 15 milligrams or less of Total Particulate per Glove.

COMMON OR USUAL NAME: **Examination Glove**

CLASSIFICATION NAME: **Examination Glove**

3. PREDICATE DEVICE IDENTIFICATION

NAME, NUMBER
a. Ultra Care Natural Rubber Latex Examination Gloves Made from Allotex® an enzyme treated natural rubber latex with a Protein Content Claim of 200 micrograms or less with 15 milligrams or less of Total Particulate per Glove. (K002718)
b. Ultra Care (GREEN) Natural Rubber Latex Examination Gloves (K982505)
c. CymunVera Powder Free GREEN Natural Rubber Latex Exmination Gloves Lined with ALOE VERA (K013286)

4. DESCRIPTION OF DEVICE

a. HOW THE DEVICE FUNCTIONS:
Natural Rubber Latex films form a barrier to body fluids and bloodborne pathogens.

b. SCIENTIFIC CONCEPTS THAT FORM THE BASIS FOR THE DEVICE:
The latex rubber is water tight under normal conditions of use. It's tensile properties cause it to conform to the hand, allowing movements necessary for a medical procedure.

c. PHYSICAL AND PERFORMANCE CHARACTERISTICS SUCH AS DESIGN,

MATERIALS
AND PHYSICAL PROPERTIES:

Natural Rubber Latex is known to create a barrier to bloodborne pathogens and and body fluids. ASTM conforming tensile properties create a glove that is strong and flexible. The leaching process removes traces of chemical accelerants that may be chemically irritating. The glove is manufactured in accordance with the requirements of ASTM D3578-00 and ASTM D5151-99 requirements.

5. STATEMENT OF INTENDED USE, INCLUDING DESCRIPTION OF THE DISEASES OR CONDITIONS THAT THE DEVICE WILL ADDRESS

This is a disposable device, intended for medical purposes, that is worn on the examiner's hand to prevent contamination between patient and examiner. Examination gloves with protein content labeling are suitable in situations where healthcare worker or patient allergic sensitivity may be a factor.

6. EXPLANATION OF SIMILARITIES OR DIFFERENCES TO PREDICATE DEVICE

The modified product has a raw material change whereby the inner surface of the glove is treated with Aloe Vera and the natural rubber latex is treated with Green pigment compared to the predicate product.

B. IF THE DECISION BASED ON PERFORMANCE DATA
1. DISCUSSION OF NON-CLINICAL TESTS

SPECIFICATION	PROPOSED	PREDICATE
	Same as predicate but with a Green color and Aloe Vera labeling claims.	**Ultra Care synthetic inner coating with no starch donning powder added with Protein and Particulate Content labeling Claims and Made from Allotex an enzyme treated natural rubber latex claim. (K002718)**
	Same Green color as predicate but with Aloe vera, Protein and particulate Content labeling Claims and Made from Allotex an enzyme treated natural rubber latex claim	**Ultra Care synthetic inner coating with no starch donning powder added and with a powder content labeling claim and GREEN colorant labeling claim. (K982505)**
	Same Aloe Vera containing as predicate but with Made from Allotex an enzyme treated natural rubber latex claim	**Cymun Vera Green synthetic inner coating with no starch donning powder added and with an Aloe Vera content labeling claim and Green colorant labeling claim. (K013286)**

PERFORMANCE STANDARDS	**ASTM D3578-00a**	**ASTM D3578-99**
WATER TIGHTNESS	**ASTM D5151-99**	**ASTM D5151-99**
RESIDUAL PROTEIN	**ASTM D5712-99**	**ASTM D5712-99**
RESIDUAL PARTICULATE	**ASTM D6124-00**	
ANTIGENIC PROTEIN	**ASTM D6499-00**	

2. DISCUSSION OF CLINICAL TESTS

SPECIFICATION	PROPOSED	PREDICATE
SAFETY		
RABBIT IRRITATION		**PASS**
GUINEA PIG SENSITIZATION		**PASS**

3. CONCLUSIONS DRAWN FROM NONCLINICAL AND CLINICAL TESTS THAT DEMONSTRATE SAFETY EFFECTIVENESS, AND PERFORMANCE =/> PREDICATE PRODUCT

The Ultra Preserve, Examination Glove has been carefully compared to legally marketed devices in the 510(k). The data summaries indicate that the proposed product meets or exceeds acceptable scores for the predicate product in nonclinical tests, and satisfies the requirements for a safe and effective, green, aloe vera, no starch donning powder added with 15 milligrams or less of total particulate with protein content labeling claim (200 micrograms or less) per glove and made from Allotex an enzyme treated natural rubber latex claim medical glove

---

Pursuant to 21 C.F.R. 807.87 (j), I, F.W. Perrella, Ph.D., Vice President R&D certify that to the best of my knowledge and belief and based upon the data and information submitted to me in the course of my responsibilities as the V.P. R&D for TILLOTSON HEALTHCARE CORPORATION, and in reliance thereupon, the data and information submitted in this of the substantial equivalence of this device have been knowingly omitted from this Submission.

F.W. Perrella, Ph.D.
Vice President R&D

DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

JUL 5 2002

Mr. Frank W. Perrella
Vice President, Research and Development
Tillotson Healthcare Corporation
360 Route 101
Bedford, New Hampshire 03110

Re: K021007
Trade/Device Name: Ultra Preserve with Aloe Vera (Green) Natural Rubber Latex Examination Gloves made from Allotex® an enzyme treated natural rubber latex with a Protein Content Labeling Claim (200 Micrograms or Less) with 15 Milligrams or Less of Total Particulate per Glove
Regulation Number: 880.6250
Regulation Name: Patient Examination Glove
Regulatory Class: I
Product Code: LYY
Dated: June 14, 2002
Received: June 14, 2002

Dear Mr. Perrella:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 21 CFR Part 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4618. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address http://www.fda.gov/cdrh/dsma/dsmamain.html

Sincerely yours,

for Timothy A. Ulatowski
Director
Division of Anesthesiology, General Hospital,
Infection Control and Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

3.0 Indications for Use Statement: Include the following or equivalent Indications for Use page. The information, data and labeling claims in the entire the 510(k) submission must support and agree with the Indications for Use statement.

INDICATIONS FOR USE

Applicant: Tillotson Healthcare Corporation

5 10(k) Number (if known):* K021007

Device Name: **Ultra Preserve With Aloe Vera (Green) Natural Rubber Latex Examination Glove Made from Allotex® an enzyme treated natural rubber latex with a Protein Content Claim of 200 micrograms or less with 15 milligrams or less of Total Particulate per Glove.**

Indications For Use:

The Ultra Preserve With Aloe Vera (Green) Examination Glove is "a disposable device intended for medical purposes that is worn on the examiner's hand to prevent contamination between patient and examiner.".

(21CFR 880.6250).

(Division Sign-Off)
Division of Dental, Infection Control,
and General Hospital Devices
510(k) Number K021007

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)
Concurrence of CDRH Office of Device Evaluation (ODE)

Prescription Use _____ OR Over-The-Counter _____
Per 21 CFR 801.109
(Optional Format 1-2-96)